



# EC-CERTIFICATE

(Full quality assurance system)



This is to certify that the company

## Peter Lazic GmbH

Immelmannweg 2  
78532 Tuttlingen  
Germany

has implemented and maintains a full quality assurance system which applies to the products at every stage from design to final controls.

Through an audit, documented in a report, performed by DQS Medizinprodukte GmbH, it was verified that the management system fulfills the requirements of

### Annex II – excluding Section 4 of Council Directive 93/42/EEC concerning medical devices

with respect to the following medical devices:

Aneurysma clip system sterile and unsterile in the variants: Class III  
Yasargil out of titanium and phynox; Perneckzy, L-  
Aneurysm-Clip and D-Clip out of titanium; each permanent  
and temporary

Neuro surgical instruments Class III

Cervical Cages Class IIb

The manufacturer is subject to surveillance according to Annex II, Section 5. The CE marking with the Notified Body Identification Number (0297) may be affixed on the devices listed in the certificate. An EC Design Examination Certificate according to Annex II, Section 4 is required for class III devices covered by this certificate. The certificate is in the case of class I(s) devices (I(s) = class I products placed on the market in sterile conditions) limited to the aspects of manufacture concerned with securing and maintaining sterile conditions. The certificate is in the case of class I(m) devices (I(m) = class I devices with a measuring function) limited to the aspects of manufacture concerned with the conformity of the products with the metrological requirements.

Certificate registration No. 004182 MR2

Certificate unique ID 170535336

Effective date 2012-03-05

Expiry date 2017-03-04

Frankfurt am Main 2012-03-05

### DQS Medizinprodukte GmbH

Frank Graichen  
Managing Director

August-Schanz-Straße 21, 60433 Frankfurt am Main, Tel. +49 (0) 69 95427-263, [medical.devices@dqs.de](mailto:medical.devices@dqs.de)

DQS Medizinprodukte GmbH is a Notified Body according to Council Directive 93/42/EEC concerning medical devices with the Identification Number 0297.

